

# Clean Test

Rapid test for the hygiene control of work surfaces



Swab for detecting protein, fat and sugar residues  
for the hygiene control of surfaces

Results within 10 minutes

Easy color change  
interpretation

Minimal training  
required



**Liofilchem**

Via Scozia - Zona Industriale  
64026 Roseto degli Abruzzi (TE) Italy  
Phone +39 0858930745  
Fax +39 0858930330  
www.liofilchem.net  
liofilchem@liofilchem.net



F52766  
Rev. 2 - 01/11/2007


# Test procedure



1  
The product consists of a plastic tube containing a swab, a vial (reagent A) and a disc (reagent B).



2  
Take the swab contained in the tube and wet the tip with 3-4 drops of physiological solution or sterile distilled water.




3  
Streak the swab on a small plot of surface measuring 10 cm x 10 cm horizontally and vertically.



4  
Insert the swab into the tube and push the tube in order to lower the swab tip close to the disc.



5  
Break the glass vial contained in the tube by pushing the tube gently.



Clean surface,  
or dirty surface after  
2 minutes reaction.



Dirty surface after  
4 minutes reaction.



Dirty surface after  
7 minutes reaction.



Dirty surface after  
10 minutes reaction.

## Clean Test

Ref. 86001

Content of the package: 50 tests



**Liofilchem**

Via Scozia - Zona Industriale  
64026 Roseto degli Abruzzi (TE) Italy  
Phone +39 0858930745  
Fax +39 0858930330  
www.liofilchem.net  
liofilchem@liofilchem.net



MICROPLANET LABORATORIOS, SL  
www.microplanet-psl.com  
info@microplanet-psl.com  
Tel. 93 590 62 92